



PSSCP Pure Steam Sample Condenser Rev0.doc Page 1 of 2

PSSCP Pure Steam Sample Condenser

PS-SC Pure Steam Sample Condenser

This PS-SC wall panel mounted is designed to allow Clean Steam (CS) and Water For Injection (WFI) samples to be taken quickly and easily whilst maintaining a sterile testing environment. This wall panel is ideal to be connect at the sampling point and can be operated with either mains or chilled water as the cooling medium. Availability of aseptic sample valve allow fine control of sample flow during testing The PS-SC can be sterilized in situ-on-line, thus ensuring continuity of samples regardless of testing frequency, ideal for fluids in pharmaceutical and purity systems applications.



Installation

Where steam is the sample medium, a valve (1) excluded from supply, suitable for continuous steam duty must be installed adjacent to steam manifold.

First of all the isolation valve on main steam (1) (or on the use point) should be connected with the valve on the top of heat exchanger (4) (tube side) by the first flexible silicone hose clamped not included in the supply; same operation for cooling medium that should be connected with the valve (5) on the lower body connection of heat exchanger (shell side) by the second flexible silicone hose clamped that which is not supplied with the wall panel unit.

The medium to be condensed/cooled passes through the tube side and typically a regulating valve (2) will be used to throttle the sample medium flow. When exit from heat exchanger, the condensate flush in the SSB sample bottle that can be disconnected and removed, when sample is finish, ready for analysis.

Cooling medium is channelled countercurrent inside the shell to ensure maximum efficiency; tipically mains or chilled water; in order to allow the flow of cooling medium to be controlled during testing an isolation valve (5) should be included. A drain valve (9) also should be included between the cooling water isolation valve (5) and heat exchanger to allow residual cooling water to be drained from the shell prior to sterilization.

The cooling water outlet has to be connected to the waste by a flexible silicone hose and must be always vented to atmosphere free of obstructions or isolation valves to prevent any build up in the shell of the heat exchanger unit.

Features:

- Aisi 316L 1.4404 stainless steel construction, suitable for PS pure steam and WFI systems
- · Self-draining design, eliminates possibility of sample retention
- Fully sterilizable in place (SIP)
- Wall panel mounted design

Typical applications:

- Pure Steam sampling
- Point of use cooling / sampling
- In-line conductivity monitoring

Operating conditions: Max working pressure

essure	: steam at 4 barg at 151°C
	: cooling water at 6 barg

@ Note – pressure rating may exceed that of clamps connections

Capacities (approximate):

- Steam: 10 kg/h of condensate at 100°C
- Water: 30 lt/h of water from 85°C to 30°C
- based on a cooling water temperature of 18°C and flow rate of 10 lt/min.

Standards:

Heat exchanger according to PED directive 97/23/CE are CE compliant under the SEP "Sound Engineering Practice" (Article 3 – Paragraph 3)



HOW TO USE

Sterilization in place (SIP)

Prior to testing or at periodic intervals it may be appropriate to sterilize the unit to ensure that sample integrity is maintained during testing.

Caution: The unit will become hot during operation and sterilization; personnel should be protected against contact with the installation with appropriate insulation.

Prior to start, to ensure that the sample bottle connection labeled "IN" is connected with the heat exchanger and connection "OUT" is connected with the steam trap.

- Open cooling water drain valve (3)(4) to allow residual cooling water to be drained from the shell. Unless drained, this water will boiling during sterilization
- 2) Open the knob labeled "STEAM" on the top of the bottle
- Slowly open the steam inlet valve (1)(2) for sterilize all surface in contact with the product ; the condensate is drained in the waste
- 4) At the completion of steaming cycle, first close the steam intercepting valve (2) secondly close the knob labeled "STEAM" on the top of the bottle.

Sampling

- A sterile sample can now be taken as follow:
- 1) Open the cooling water inlet valve (5) and ensure that a flow is present
- 2) Open the knob labeled "PRODUCT" on the top of the bottle
- 3) Open the steam inlet valve (1)(2)
- 4) Once a suitable sample has been obtained, close the knob "PRODUCT"
- 5) Close the steam inlet valve (1)(2)
- 6) Cooling medium should be allowed to flow for a short period in order to condense/cool any medium passing through the LSC. When no further sample is evident and the system has cooled, isolate the cooling water supply valve (5)
- 7) Open the knob labeled "STEAM" on the top of the bottle to allow any residual heat energy to be dissipate and the entire system drained
- 8) Close the knob labeled "STEAM" on the top of the bottle
- 9) Disconnect the sample bottle by the clamp connections ready for laboratory analysis

Product Specification

Materials:

Tube, valves and fittings in AISI 316L – 1.4404 flexible hose not included clamp gaskets in PTFE

Dimensions and weight: 800 x 400 x h max. 240 mm

empty weight 000 kg. approx

Connections: Steam: ½" clamp BS 4825 Cooling water: ¾" clamp BS 4825

Supply included:

All valves and parts showed in picture / sketch

Available accessories:

flexible hose Silicone Platinum Cured APSW-P for connection





Get the information you need and more at : info@aerreinox.it

In the interests of development and improvement of the product, we reserve the right to change the specifications without prior notice.

